



GOVERNMENT OF JAMMU AND KASHMIR
HEALTH AND MEDICAL EDUCATION DEPARTMENT

GOVERNMENT MEDICAL COLLEGE, ANANTNAG

Email ID: gmcanantnag2018@gmail.com

PURCHASE Section

Website: www.gmcanantnag.net

No: GMCA/Pur/ 279/140-41

Dated: 23-04-2025


Quotation

Quotations in a sealed envelope are invited from interested firms registered with appropriate Govt. Authority for the supply of **Infusion pumps for Cardiology Section /SICU** of MMABM Associated Hospital, GMC Anantnag.

Item Name	Specification	Quantity
Infusion pumps	Annexure A	12 No

Note: -

1. The items will be procured **Directly through GeM portal** from those Firms/Agencies/Bidders only who will submit their Rate Quotations within the stipulated time frame.
2. The Bidders shall quote the rates of only those items, which they can Supply through **GeM Portal**
3. The quotations are required just to check the reasonability of rates, before procuring the said items through **GeM-Portal**
4. The Quotations in a sealed envelope shall reach this office by or before **28-04-2025** till 12:00 noon
5. **No compromise in quality.**
6. The Sealed quotations should have the company seal and signature with stamp duty.


Principal/Dean
22/4/25
Government Medical College,
Anantnag.

Copy for information to the: -

1. Chief Accounts Officer, GMC, Anantnag.
2. I/C website for information with the directions to upload the quotation on college website for wider publicity.

Documents to be enclosed:

The bidder must upload **Authorization certificate** of the similar item.

GST Registration Certificate.

The bidders are directed to Enclose the **catalogue** of the item with mentioned specifications

The bidders may be asked to bring the sample/demo which shall be readily available.

No compromise in specification and quality.





GOVERNMENT OF JAMMU AND KASHMIR
HEALTH AND MEDICAL EDUCATION DEPARTMENT

GOVERNMENT MEDICAL COLLEGE, ANANTNAG

Email ID: gmc.anantnag2018@gmail.com

PURCHASE Section

Website: www.gmc.anantnag.net

ANNEXURE 'A'

Syringe Infusion Pump

- *Portability:* Lightweight and portable, suitable for hospital and ambulatory use

Technical Specifications:

Syringe Compatibility

- Accepts a wide range of syringe sizes (typically 1 mL to 60 mL)
- Automatic syringe recognition (depending on model version)

-Infusion Modes:

- Continuous infusion
- Intermittent (bolus) delivery
- Patient-controlled analgesia (PCA) mode (if supported)

Flow Rate Range:

- Typically *0.1 mL/h to 999 mL/h* (varies by syringe size)

- Dose Calculation:

- Supports weight-based dosing (mg/kg/h)

- Accuracy:

- $\pm 2\%$ (meets international infusion standards)

- Occlusion Pressure:

- Adjustable occlusion alarm settings (e.g., high/medium/low)



GOVERNMENT OF JAMMU AND KASHMIR
HEALTH AND MEDICAL EDUCATION DEPARTMENT

GOVERNMENT MEDICAL COLLEGE, ANANTNAG

Email ID: gmcanantnag2018@gmail.com

PURCHASE Section

Website: www.gmcanantnag.net

Battery:

- Rechargeable battery with 6+ hours of operation (varies by usage)
- Battery status indicator

Display & Alarms:

- Display: Backlit LCD screen
- Alarms:
 - Occlusion detection
 - End of infusion
 - Low battery
 - Syringe misload
 - Air-in-line detection (if equipped)

Physical Specifications:

- **Dimensions: 200 x 100 x 150 mm (approximate)
- Weight: 1.5 kg (varies slightly by model)

Dr. Anant

[Signature]

Anant
22/4/25
Principal/Dean
Government Medical College,
Anantnag.